The application of extended modified Lambert Beer model for measurement of blood carboxyhemoglobin and oxyhemoglobin saturation

Audrey Huong*

*Department of Electronic Engineering
Faculty of Electrical and Electronic Engineering
University Tun Hussein Onn Malaysia
Batu Pahat 86400, Johor, Malaysia
audrey@uthm.edu.my*

Xavier Ngu

*Department of Communication Engineering
Faculty of Electrical and Electronic Engineering
University Tun Hussein Onn Malaysia
Batu Pahat 86400, Johor, Malaysia*

Received 16 August 2013

Accepted 6 December 2013

Published 5 February 2014

This work presents the use of extended Modified Lambert Beer (MLB) model for accurate and continuous monitoring of percent blood carboxyhemoglobin (COHb) (SCO) and oxyhemoglobin (OxyHb) saturation (SO₂) via a fitting procedure. This quantification technique is based on the absorption characteristics of hemoglobin derivatives in the wavelength range of 520–600 nm to give the best estimates of the required parameters. A comparison of the performance of the developed model and MLB law is made using attenuation data from Monte Carlo simulations for a two-layered skin model. The results revealed a lower mean absolute error of 0.4% in the values estimated by the developed model as compared to 10% that is given by the MLB law. This study showed that the discussed approach is able to provide consistent and accurate measurement of blood SO₂ and SCO across different skin pigmentations suggesting that it may potentially be used as an alternative means for clinical diagnosis of carbon monoxide (CO) poisoning.

Keywords: Blood carboxyhemoglobin saturation; carbon monoxide poisoning; blood oxyhemoglobin saturation; modified Lambert Beer law.

*Corresponding author.

This is an Open Access article published by World Scientific Publishing Company. It is distributed under the terms of the Creative Commons Attribution 3.0 (CC-BY) License. Further distribution of this work is permitted, provided the original work is properly cited.

1. Introduction

The concentration of carbon monoxide (CO) in blood or carboxyhemoglobin (COHb) is the key parameter used to evaluate one's CO toxicity level after exposure to the flammable, toxic gas. The elevated level percent of COHb in blood to 60% can prove fatal¹ as CO molecules, which have a higher tendency to bind with hemoglobin, reduce the concentration of oxygen bound to hemoglobin (known as oxyhemoglobin (OxyHb)). The accurate measurement of both of these parameters is, therefore, essential so that suitable oxygen therapy can be administered in response to the different severity of CO poisoning. Nonetheless the conventional technique used for accurate measurement of percent blood COHb saturation (SCO) is limited to those that involve collecting and analyzing blood sample, which is impractical for continuous monitoring of one's blood CO level. Non-invasive and continuous measurement of COHb can be provided by pulse CO-oximeter but this device is shown to produce large error in the estimated value of 14%.² Meanwhile, fast and non-invasive measurement of arterial blood OxyHb saturation (SO₂) can be provided by pulse oximetry, but its high accuracy is limited to SO₂ greater than 70%. Techniques that are able to provide fast and accurate measurement of percent saturation of oxygen and CO in blood are of significant value, especially to individuals who are occupationally exposed to incomplete combustion of organic materials such as that of coal-fired power plant operators and firefighters.

Although the absolute concentration measurement of an absorber using signals from continuous wave (CW) spectroscopy is only feasible if the path-length of photons travelled in the medium is known,³ for example, using the measurement of temporal point spread function of light traversing a scattering medium, the percentage of hemoglobin components can be deduced using light attenuation measured from CW spectroscopy performed on skin.⁴ The use of CW spectroscopic technique allows the non-invasive measurement of fractional concentration of an absorbing component by means of fitting using either analytic models or a library of data simulated using Monte Carlo method or diffusion model.^{4,5} Amongst the commonly used analytic models is Modified Lambert Beer (MLB) law that assumes a linear relationship between light absorption and attenuation. In addition to the fitting method, the required

parameter is also commonly solved via simultaneous solution of this linear model. The relationship between light attenuation and absorption in blood is, however, a nonlinear and multi-valued function owing to the wavelength dependent scattering properties of different skin layers. This results in poor accuracy of the value estimated by the MLB. Other more complex models such as power law model developed from radiative transfer equation and cubic attenuation model are designed to provide a more accurate estimate of this value.^{4,6} However, their application is limited. Besides, because of the high melanin absorption and high scattering of skin tissue at the visible wavelengths, many of these workers preferred working with signals in the near infrared range.

This paper presents the use of an extended MLB model to extract the percentage saturation of COHb and OxyHb via a fitting algorithm. The performance of this model is evaluated using data given by Monte Carlo simulations as actual fractional concentrations of absorbers are known. The focus of this study is on signals measured across visible range as this gives greater flexibility in the selection of equipment during experiment, and this work aims to estimate both blood SCO and SO₂ non-invasively using data collected from optical spectroscopy performed on skin surface without attempting to solve fractional concentration of deoxyhemoglobin (Hb).

2. Methods

2.1. Mathematical model of skin and light propagation simulations

Skin is a complex structure with inhomogeneous distribution of tissues, capillaries and blood vessels. Many previous studies have divided skin into three distinctive layers, namely, epidermis, dermis and subcutaneous layer. Recent efforts to obtain a closer approximation to the actual skin include modeling of skin as a nine layer medium having different optical properties. In this work, a two-layered skin model similar to that used in Yudovsky and Pilon⁷ is established, wherein the upper layer is an infinite slab with a commonly reported thickness of 100 μm ⁸ supported by a semi-infinite bottom layer. These upper and bottom layers are taken to be bloodless epidermis and blood-rich dermis layers, respectively, and are modeled to have the optical properties of the corresponding medium. Subcutaneous layer has not been considered in this skin model owing to the limit

of light penetration depth of approximately 2 mm in skin.⁴ Even though the medium's refractive index is a wavelength dependent function, this variation is small in the considered visible wavelength range,⁹ and so is assumed as $n = 1.4$ for both layers.

Since the thickness of the epidermis is considerably lesser than that of the dermis, both of these layers have the same wavelength dependent reduced scattering coefficients. This is a result of the negligible differences observed on the simulated data when other scattering coefficients are used for epidermis layer.¹⁰ The wavelength dependent μ'_s values used in this work in Fig. 1 are taken from the results of Staveren *et al.*¹¹ on intralipid 10% as the reported μ'_s range and monotonic variation in μ'_s with wavelength are similar to that of biological tissues presented in other works.¹² Light absorption by melanin in the assumed blood-free epidermis, $\mu_{a_{\text{epi}}}$, is taken from Meglinski and Matcher,⁸ wherein changes in $\mu_{a_{\text{epi}}}$ with wavelength, λ (in nanometers), can be approximated by⁸:

$$\mu_{a_{\text{epi}}} = 5 \times 10^9 \lambda^{-3.33} (\text{mm}^{-1}). \quad (1)$$

The epidermal absorption coefficient, $\mu_{a_{\text{epi}}}$, is given by the product of $\mu_{a_{\text{mel}}}$ times volume fraction of melanosomes in epidermis. The volume fraction of melanosomes, $f_{v\text{-mel}}$, ranges from 1–3% for a fair skinned person and up to 40% for darkly pigmented skin.¹³ The $\mu_{a_{\text{epi}}}$ values given from $\mu_{a_{\text{mel}}}$ in Eq. (1) and $f_{v\text{-mel}}$ of 40% chosen here for demonstration purposes are shown in Fig. 1.

Fig. 1. Reduced scattering coefficients, μ'_s , and epidermal absorption coefficients ($\mu_{a_{\text{epi}}}$), absorption coefficients of OxyHb ($\mu_{a_{\text{OxyHb}}}$), Hb ($\mu_{a_{\text{Hb}}}$) and COHb ($\mu_{a_{\text{COHb}}}$). Total hemoglobin concentration is taken to be $268 \mu \cdot \text{mol} \cdot \text{L}^{-1}$.

Capillaries and blood vessels carrying hemoglobin molecules absorb light in the dermis layer, with blood OxyHb, Hb and COHb taken here as three absorbers in this medium. The total light absorption in the dermis, $\mu_{a_b}(\lambda)$, is given by the summation of the product of concentration, C , and wavelength dependent extinction coefficient, $\varepsilon(\lambda)$, of each absorbing species (indicated by subscript notation) as follows:

$$\mu_{a_b}(\lambda) = \varepsilon_{\text{OxyHb}}(\lambda)C_{\text{OxyHb}} + \varepsilon_{\text{Hb}}(\lambda)C_{\text{Hb}} + \varepsilon_{\text{COHb}}(\lambda)C_{\text{COHb}}. \quad (2)$$

Given that $T = C_{\text{OxyHb}} + C_{\text{Hb}} + C_{\text{COHb}}$ is the total hemoglobin concentration, Eq. (2) can be written as Eq. (3). Next, rearranging Eq. (3) gives Eq. (4), wherein SO_2 and SCO in Eq. (4) are blood OxyHb and COHb saturation, respectively, given by:

$$\text{SO}_2 = \frac{C_{\text{OxyHb}}}{T},$$

$$\text{SCO} = \frac{C_{\text{COHb}}}{T},$$

$$\mu_{a_b}(\lambda) = \varepsilon_{\text{OxyHb}}(\lambda)C_{\text{OxyHb}} + \varepsilon_{\text{COHb}}(\lambda)C_{\text{COHb}} + \varepsilon_{\text{Hb}}(\lambda)(T - C_{\text{OxyHb}} - C_{\text{COHb}}), \quad (3)$$

$$\mu_{a_b}(\lambda) = ((\varepsilon_{\text{OxyHb}}(\lambda) - \varepsilon_{\text{Hb}}(\lambda))\text{SO}_2 + (\varepsilon_{\text{COHb}}(\lambda) - \varepsilon_{\text{Hb}}(\lambda))\text{SCO} + \varepsilon_{\text{Hb}}(\lambda))T. \quad (4)$$

Using the total hemoglobin concentration of $T = 268 \mu \cdot \text{mol} \cdot \text{L}^{-1}$, which is calculated from the reported dermis blood volume fraction of 5%,¹³ and the extinction coefficient of different hemoglobin derivatives reported by Zijlstra,¹⁴ the absorption of each absorber in dermis layer is shown in Fig. 1.

The distinctive differences in the absorption of hemoglobin components can be observed in the wavelength range of 520–600 nm in Fig. 1. The quantification approach employed in this study depends on this absorption spectra and a fitting procedure that will be described in Sec. 2.2 to give the best guess of the required parameters, so this paper works with signals measured across this wavelength range to find the value of SO_2 and SCO .

This work used attenuation data generated using Monte Carlo simulation code described in Chang *et al.*¹⁵ to verify the accuracy of value estimated by the developed analytic model and to compare the performance of the corresponding model with that

of the MLB law. For each μ'_s value in the wavelength range of 520–600 nm in Fig. 1, the behavior of light propagate in the medium is simulated after 1×10^6 photons are launched normally into the medium. Photons propagate with optical path-length greater than 100 mm are discarded owing to their low, if any, contribution to the overall measured intensity. Since the skin is modeled in this work as an infinitely thick medium, the simulation data is derived from reflectance mode measurement. The detection of photons is via an annular detector with inner and outer detection radius of 2 mm and 4 mm, respectively, from the light source. This detector is placed directly at the skin surface.

For demonstration purposes, 16 arbitrarily selected combinations of percent SO_2 and SCO , c_x , used to give dermis total absorption coefficient shown in Eq. (4) are tabulated in Table 1, where the subscript x represents the combination index. In addition, an investigation of the consistency in the performance of the developed model using data measured from populations with different skin pigmentation is by considering $f_{\text{v-mel}}$, of 2% and 40% in this study. The effects of these dermal and epidermal absorption are incorporated post simulation into Monte Carlo simulated data according to Beer's law to speed up the simulation process.

2.2. Extended modified Lambert Beer model and fitting procedure

The common practice employed in solving an absorber's fractional concentration value is by the simultaneous solution of MLB in Eq. (5) or via the fitting of data simulated from light propagation model to the measured spectra.^{5,16} This paper proposes the use of an attenuation model extended from MLB model (hence the name extended MLB) shown in Eq. (6) to find the value of SO_2 and SCO in Eq. (4) by means of fitting to the measured spectra

$$A(\lambda) = G_0 + \mu_a d_0, \quad (5)$$

$$A(\lambda) = G_0 + \mu_a d_0 + G_1 \lambda + \lambda \exp(-\mu_a d_1). \quad (6)$$

Here μ_a represents light absorption. The first and second terms in Eq. (6) are from the MLB describing the light attenuation offset and absorption dependent light attenuation, respectively, with the parameter d_0 normally taken to be "light path-length".

Since both scattering coefficients and light absorption by epidermis decrease approximately linearly with wavelength in the considered wavelength range in Fig. 1, this work assumed light attenuation due to scattering process and light absorption in the epidermis is represented by parameter $G_1 \lambda$ in Eq. (6) so that μ_a in second and last term of this equation is given from the absorption by blood rich dermis expressed in Eq. (4). This gives an expression of light attenuation as a function of blood OxyHb and COHb saturation in Eq. (7) with the symbol $\Delta \varepsilon_1$ and $\Delta \varepsilon_2$ in Eq. (7) is given by $\varepsilon_{\text{OxyHb}}(\lambda) - \varepsilon_{\text{Hb}}(\lambda)$ and $\varepsilon_{\text{COHb}}(\lambda) - \varepsilon_{\text{Hb}}(\lambda)$, respectively. Meanwhile the exponential term is included in this model to represent the light attenuation as a complex function of dermal light scattering, which is assumed to yield changes in "light path-length", d_1 , and absorption in the dermis.

The value of SCO and SO_2 is found by fitting this model to the simulated data using *fminsearch* function in MATLAB. This optimization method does not apply constraints to the value of these fitting parameters but it will seek for a new value of these parameters according to the size of error between the value given from the employed analytic model and the real attenuation, ΔE , during each iteration. This fitting process is terminated when either the mean absolute ΔE is less than 1×10^{-20} or the iteration process has exceeded 1000, when the optimal value of SCO and SO_2 is assumed to have been obtained

$$A(\lambda) = G_0 + (\Delta \varepsilon_1 \text{SO}_2 + \Delta \varepsilon_2 \text{SCO} + \varepsilon_{\text{Hb}}(\lambda)) d_0 T + G_1 \lambda + \lambda \exp(-(\Delta \varepsilon_1 \text{SO}_2 + \Delta \varepsilon_2 \text{SCO} + \varepsilon_{\text{Hb}}(\lambda)) d_1 T). \quad (7)$$

Table 1. Different combinations of percent blood OxyHb (SO_2) and COHb saturation (SCO), c_x , to give dermis absorption coefficients used in post Monte Carlo stage.

c_x	c_1	c_2	c_3	c_4	c_5	c_6	c_7	c_8	c_9	c_{10}	c_{11}	c_{12}	c_{13}	c_{14}	c_{15}	c_{16}
SO_2 (%)	100	95	95	90	90	85	85	80	80	75	70	65	60	55	50	45
SCO (%)	0	3	5	5	10	10	15	10	20	25	30	35	40	45	50	55

3. Results

The Monte Carlo simulated reflectance data for a two-layered medium with optical properties described in Sec. 2.1 are used here to compare the accuracy of percent OxyHb and COHb saturation given by MLB and extended MLB. The estimation of SO_2 and SCO value using MLB is presented here first. The estimation begins by substituting μ_{a_b} in Eq. (4) into μ_a in Eq. (5) to give an expression of attenuation as a

linear function of SO_2 and SCO. The fitting procedure described in Sec. 2.2 is then used to search for the best estimates of these parameters using the MLB. The difference between the actual and estimated percent SO_2 and SCO value represented by ΔSO_2 and ΔSCO , respectively, is defined as:

$$\Delta\text{SO}_2 = \text{SO}_{2\text{EST}} - \text{SO}_{2\text{ACT}}, \quad (8)$$

$$\Delta\text{SCO} = \text{SCO}_{\text{EST}} - \text{SCO}_{\text{ACT}}, \quad (9)$$

Fig. 2. Error in blood OxyHb saturation (ΔSO_2) and blood COHb saturation (ΔSCO) given by MLB for different combinations of SO_2 and SCO in Table 1 and when volume fraction of melanosomes ($f_{v\text{-mel}}$) is (a) 2% and (b) 40%.

Fig. 3. Error in blood OxyHb saturation ΔSO_2 (“o”) and blood COHb saturation ΔSCO (“+”) estimated by extended MLB when volume fraction of melanosomes ($f_{v\text{-mel}}$) is (a) 2% and (b) 40%. The actual SO_2 and SCO used in Monte Carlo simulations are listed in Table 1.

where SO_{2_EST} and SCO_{EST} indicate that of the value estimated by the employed model while the actual percent SO_2 and SCO value listed in Table 1 are symbolized by SO_{2_ACT} and SCO_{ACT} . The ΔSO_2 and ΔSCO given from the MLB against different combinations of SO_{2_ACT} and SCO_{ACT} in Table 1 and at f_{v_mel} of 2% and 40% are represented by symbol “o” and “+”, respectively, and plotted in Figs. 2(a) and 2(b).

An investigation of the accuracy of the estimates using extended MLB as compared to that of MLB is by using the same reflectance data and fitting method that are used to test the performance of the MLB in Fig. 2. The best guess of percent SO_2 and SCO at different combinations in Table 1 given from extended MLB in Eq. (7) is then used to calculate ΔSO_2 and ΔSCO from Eqs. (8) and (9), and are plotted in Fig. 3.

4. Discussion

The comparison between the accuracy of values estimated by MLB and extended MLB in Figs. 2 and 3 shows a higher error in the values given by the MLB. The calculated mean absolute error is approximately 10% as compared to 0.4% that is produced by the extended MLB. This is likely due to the linear and single-valued light attenuation and blood absorption relationship that is assumed by the MLB. Furthermore, light absorption by melanin has not been considered in this model.

The error in the SCO and SO_2 value determined by MLB in Fig. 2 is, in general, decreased with the combination index shown in Table 1. Besides the result also showed the consistent overestimation of SCO and underestimation of SO_2 value at different f_{v_mel} and combination index. An investigation of the attenuation spectrum of these SO_2 and SCO combinations reveals that two hemoglobin peaks in the considered wavelength range of 520–600 nm, which is attributed to the unique absorption features of OxyHb in Fig. 1, are apparent when $SO_2 = 100\%$ and $SCO = 0\%$. These peaks gradually diminished with a decrease in SO_2 (or an increase in SCO). A possible explanation of the observed error pattern is that the MLB fails to sufficiently describe light attenuation contributed by underlying absorption (by different absorbers) and scattering processes, producing local minimum of the function that is not globally optimal during the fitting process when the hemoglobin peaks

are at their highest. Another interesting fact that can be observed in Fig. 2 is the inconsistent magnitude of ΔSCO given by the MLB as the f_{v_mel} varied. This could be due to the reason that an increase in epidermal light absorption reduces the hemoglobin peaks, which has the similar effect to that when the percent blood COHb saturation increases, leading to an inconsistent error in SCO given by the MLB for measurement on different pigmented skin.

The considerably consistent error in the SO_2 and SCO value estimated by extended MLB across different volume fraction of melanosomes in Fig. 3 with the mean of their differences in absolute ΔSO_2 and ΔSCO given by 0.03% and 0.056%, respectively, implies that the effects of epidermal melanin absorption on the measured attenuation are reasonably accurately represented by the wavelength dependent linear term in the developed model in Eq. (7). Even though the exponential term in this model is to account for inseparable effects of light scattering and absorption in the dermis on the measured signals, there is an overall slight underestimate of SO_2 and SCO values in Figs. 3(a) and 3(b) independent of f_{v_mel} value. This could be resulted from the insufficiency of this model at describing light propagation behavior in the dermis layer of the skin model. These errors are, however, considerably small ($< \pm 1\%$), suggesting the ability of this model to resolve the light attenuation contributed by the different absorption of the present absorbers. Since this study worked with signals in the wavelength range of 520–600 nm, wherein the variation in absorption characteristics of dermal absorbers in Fig. 1 is most distinctive, the low errors shown in Fig. 3 confirm the suitability of the corresponding wavelengths for quantification works using the developed extended MLB. It should be noted that the use of a narrower wavelength range such as that of 520–580 nm which contains the corresponding absorption spectra variation will produce the same result.

The viability of using extended MLB and the described strategy to determine blood SO_2 and SCO with reasonable accuracy indicates that this can be a suitable means for fast and continuous monitoring of one's percentage OxyHb and COHb, and *in vivo* experiment works are currently underway to confirm this. The technique described here does not depend on the calibrated data as is required in pulse oximetry and would be insensitive to noise on signals in the measurement of SO_2 and

SCO. Furthermore, the effects of light absorption by melanin has been accounted for in this developed model, hence accuracy of the measurements would not be compromised by differences in skin color. This renders the discussed fitting technique and model potentially to be used particularly for CO poisoning diagnosis and treatment, or as a cost effective health care tool used on power plant workers and firefighters who are at risk of flammable and toxic gas exposure.

5. Conclusion

This study demonstrates the use of extended MLB law model for non-invasive and continuous monitoring of percentage saturation of oxygen and CO in blood. This work concludes that accurate measurement of these values can be obtained if the developed fitting model has sufficiently described the true underlying light absorption and scattering processes. In addition, wavelengths in the range of 520–600 nm is shown to be suitable as working wavelengths for the estimation of blood SO₂ and SCO using CW spectroscopic measurement data.

References

1. A. L. Ilano, T. A. Raffin, Kienle, "Management of carbon monoxide poisoning," *Chest* **97**, 165–169 (1990).
2. M. Touger, A. Birnbaum, J. Wang, K. Chou, D. Pearson, P. Bijur, "Performance of the RAD-57 Pulse Co-oximeter compared with standard laboratory carboxyhemoglobin measurement," *Am. College of Emergency Phys.* **56**(4), 382–388 (2010).
3. G. Zhang, A. Katz, R. R. Alfano, A. D. Kofinas, D. A. Kofinas, P. G. Stubblefield, W. Rosenfeld, D. Beyer, D. Maulik, M. R. Stankovic, "Brain perfusion monitoring with frequency-domain and continuous-wave near-infrared spectroscopy: A cross-correlation study in newborn piglets," *Phys. Med. Biol.* **45**, 3143–3158 (2000).
4. A. Vogel, V. V. Chernomordik, J. D. Riley, M. Hassan, F. Amyot, B. Dasgeb, S. G. Demos, R. Pursley, R. F. Little, R. Yarchoan, Y. Tao, A. H. Gandjbakhche, "Using noninvasive multispectral imaging to quantitatively assess tissue vasculature," *J. Biomed. Opt.* **12**(5), 051604 (2007).
5. A. Pifferi, P. Taroni, G. Valentini, S. Andersson-Engels, "Real-time method for fitting time-resolved reflectance and transmittance measurements with a Monte Carlo Model," *Appl. Opt.* **37**(13), 2774–2780 (1998).
6. M. Kobayashi, Y. Ito, N. Sakauchi, I. Oda, I. Konishi, Y. Tsunazawa, "Analysis of nonlinear relation for skin hemoglobin imaging," *Opt. Exp.* **9**(13), 802–812, (2001).
7. D. Yudovsky, L. Pilon, "Rapid and accurate estimation of blood saturation, melanin content, and epidermis thickness from spectral diffuse reflectance," *Appl. Opt.* **49**(10), 1707–1719 (2010).
8. I. V. Meglinski, S. J. Matcher, "Quantitative assessment of skin layers absorption and skin reflectance spectra simulation in the visible and near-infrared spectral regions," *Physiol. Measurement* **23**(4), 741–753 (2002).
9. T. Maeda, N. Arakawa, M. Takahashi, Y. Aizu, "Monte Carlo simulation of spectral reflectance using a multilayered skin tissue model," *Opt. Rev.* **17**(3), 223–229 (2010).
10. A. Kienle, R. Hibst, "A new optimal wavelength for treatment of port wine stains?" *Phys. Med. Biol.* **40**(10), 1559–1576 (1996).
11. H. J. Van Staveren, C. J. M. Moes, J. Van Marle, M. J. C. Van Gemert, S. A. Prahl, "Light scattering in Intralipid-10 in the wavelength range of 400–1100 nm," *Appl. Opt.* **30**, 4507–4514 (1991).
12. A. N. Bashkatov, E. A. Genina, V. I. Kochubey, V. V. Tuchin, "Optical properties of human skin, subcutaneous and mucous tissues in the wavelength range from 400 to 2000 nm," *J. Phys. Appl. Phys.* **38**, 2543–2555 (2005).
13. S. L. Jacques, D. J. McAuliffe, "The melasome: Threshold temperature for explosive vaporization and internal absorption coefficient during pulsed laser irradiation," *Photochem. Photobiol.* **53**, 769–775 (1991).
14. W. G. Zijlstra, A. Buursma, O. W. Van Assendelft, *Visible and Near Infrared Absorption Spectra of Human and Animal Haemoglobin: Determination and Application*, VSP International Science Publisher, Netherlands (2000), pp. 180–194.
15. P. C. Y. Chang, J. G. Walker, K. I. Hopcraft, B. Ablitt, E. Jakeman, "Polarization discrimination for active imaging in scattering media," *Opt. Commun.* **159**, 1–6 (1999).
16. R. N. Pittman, B. R. Duling, "A new method for the measurement of percent oxyhemoglobin," *J. Appl. Physiol.* **38**(2), 315–320 (1975).