A review of microwave-induced thermoacoustic imaging: Excitation source, data acquisition system and biomedical applications

Yongsheng Cui, Chang Yuan and Zhong Ji*

*MOE Key Laboratory of Laser Life Science and Institute of Laser Life Science
College of Biophotonics, South China Normal University
Guangzhou 510631, P. R. China
jizhong@scnu.edu.cn

Received 29 March 2017

Accepted 24 May 2017

Published 27 June 2017

Microwave-induced thermoacoustic imaging (TAI) is a noninvasive modality based on the differences in microwave absorption of various biological tissues. TAI has been extensively researched in recent years, and several studies have revealed that TAI possesses advantages such as high resolution, high contrast, high imaging depth and fast imaging speed. In this paper, we reviewed the development of the TAI technique, its excitation source, data acquisition system and biomedical applications. It is believed that TAI has great potential applications in biomedical research and clinical study.

Keywords: Thermoacoustic imaging; microwave; biomedical applications.

1. Introduction

Medical imaging is a necessary tool in medical practice. Among them, X-ray imaging including CT is the clinical gold standard technique to detect abnormal structure. Nonetheless, the ionizing radiation damage limits its applications. Other imaging techniques (such as magnetic resonance imaging and ultrasound imaging) suffer from low specificity or low contrast in images and are only sensitive for selected populations. Optical imaging has high resolution and high contrast, but it suffers from low

imaging depth. Therefore, it is urgent to develop a new technique that can surmount these difficulties.

As an emerging medical imaging method, microwave-induced thermoacoustic imaging (TAI), which combines the advantages of both high imaging contrast from electromagnetic absorption and high resolution of ultrasound imaging, can achieve submillimeter resolution at decimeter depth. Kruger *et al.*^{1–3} and Wang *et al.*^{4–27} open up this field almost at the same time. With the publication of above reports, more and more researchers are aware of the advantages of TAI and put their efforts into

*Corresponding author.

improving TAI system. Nowadays, the main TAI groups in this field are Patch *et al.*,^{28–33} Ntzia-christos *et al.*,^{34–37} Jiang *et al.*,^{38–43} Xin *et al.*,^{44–49} Arbabian *et al.*,^{50,51} Zheng *et al.*,^{52–54} Xing *et al.*^{80–98} and others.^{55–79} Under the efforts of these researchers, the excitation source, data acquisition system and image algorithm have been improved greatly, and it can be revealed that TAI is efficient in detecting the area with anomalous microwave absorption in biological tissue, particularly for breast cancer. This paper will review and discuss the TAI system briefly and focus on the introduction of excitation source, data acquisition system and biomedical applications.

2. The Principle

When a microwave pulse irradiates to biological tissue, a small temperature rise can be produced, and the heated tissue will expand and shrink into an acoustic source. When microwave energy delivered in a pulse is short enough so that thermal diffusion can be ignored during the pulse, the resulting thermoacoustic pressure $P(r, t)$ that reaches a detector at position r and time t can be expressed as the following equation^{1–3}:

$$P(r, t) \approx \frac{\beta I_0 C_0}{4\pi C_p} \tau \frac{d}{dt} \oint_{|r-r'|=C_0 t} A(r') \frac{dr'}{C_0 t}, \quad (1)$$

where β is the isobaric volume expansion coefficient, I_0 a factor proportional to the incident optical energy density, C_0 the speed of ultrasound, C_p the specific heat, τ the microwave pulse width and $A(r')$ the absorbed energy density per unit volume of soft tissue at position r' at time t . The induced thermoacoustic wave propagates from its originating region and can be detected by an ultrasonic detector. The strength and the profile of a thermoacoustic signal depend on the microwave absorption of the tissue.

Although sharing the similar principle with the photoacoustic imaging (PAI), TAI may have broader applications in medical imaging. The effect of microwave on biological tissue is mainly reflected in the electric field, and electrical field at the depth z can be expressed as:

$$E = E_0 \exp(-\alpha z), \quad (2)$$

where E is the electrical field at depth z , E_0 the electrical field at the sample surface and α the

electric field absorption coefficient expressed as:

$$\alpha = \omega \sqrt{\frac{\mu \varepsilon}{2} \left(\sqrt{1 + \left(\frac{\sigma}{\omega \varepsilon} \right)^2} - 1 \right)}, \quad (3)$$

where ω is the angular frequency, μ the permeability, ε the permittivity and σ the conductivity. The induced acoustic pressure is expected to depend on the intensity of microwave and the dielectric constant of the material. In the frequency range from 0.1 to 10 GHz, the relative dielectric constant has a value of 5–70 for soft tissues, and the conductivity has a value of 0.02–3 S/m. The dielectric properties of tissue determine the absorption of microwave at various frequencies. For example, at 9.4 GHz, the penetration depth, which is the inverse of the absorption coefficient, is 3.4 and 0.35 cm for fat and muscle; whereas at 500 MHz, the penetration depths for fat and muscle are 23.5 and 3.4 cm. Most of the other soft tissues have an absorption coefficient between muscle and fat. Moreover, as long as microwave energy reaches the target, the thermoacoustic effect will be generated. Therefore, TAI may potentially be used to detect the deep object with anomalous microwave absorption.

3. The Excitation Source

TAI and PAI technology are based on the thermoelastic effect, and the main difference is the excitation source. TAI uses microwave as the excitation source, and PAI uses laser. Because of high penetration depth of microwave, TAI obtains a good imaging depth, generally at 6–15 cm. However, the microwave beam is hard to focus due to diffraction, so that the existing excitation sources have high-power outputs, mainly divided into three types, namely, magnetron modulation mode, high-voltage discharged mode and modulated continuous mode.

3.1. Magnetron modulation mode

Microwave generator in magnetron modulation mode is usually composed of electron emitter, power supply, magnetron and waveguide, as shown in Fig. 1. The electrons from electron emitter can be accelerated and generate stable electronic oscillation in magnetron. Then the electronic oscillation can be transferred to microwave in the waveguide. At the end of waveguide, microwave transforms from waveguide pattern to radiation pattern due to

Fig. 1. Schematic diagram of magnetron modulation mode.

diffraction. As oscillation in magnetron is stable and the waveguide has a frequency-selective effect, microwave in this situation has a single frequency and be suitable for mapping absorption coefficient. At the meantime, sample under test is usually placed at the end of the waveguide, where energy is relatively concentrated, so that a better signal-to-noise ratio (SNR) can be obtained. Therefore, most researchers including Wang *et al.*, Jiang *et al.*, Xin *et al.*, Patch *et al.*, Ntziachristos *et al.* and Xing *et al.* adopted waveguide mode. Details can be seen in Table 1. Figure 2(a) shows a typical microwave generator working in the waveguide mode and Fig. 2(b) depicts several pulse widths of this microwave generator.

However, those microwave generators have two weaknesses. One is that the main frequency cannot change due to the fixed operating frequency of magnetron and waveguide, i.e., against selective excitation. The other is that the pulse width is difficult to narrow due to manufacturing technology of magnetron, and the minimum is commonly

0.2–0.3 μs , which limits imaging resolution to about 0.5 mm.

3.2. High-voltage discharged mode

In order to get a narrower pulse width and to improve imaging resolution, microwave in high-voltage discharged mode has been adopted. As is shown in Fig. 3, the microwave generator is usually composed of oscillation circuit, boosting transformer, high-speed switch and antenna. Similar to the magnetron, the electrons can form electrical pulses in the oscillation circuit and then boost to high voltage by transformer. Finally, the high-voltage pulse propagates through the antenna into unbounded medium, namely, microwave. The pulse width of microwave can reach a few nanoseconds depending on the speed of the switch, so that the resolution can realize about dozens of microns' range. Ntziachristos *et al.* and Xing *et al.* adopted microwave in this mode, and the details can be seen in Table 1.

Especially, the ultrashort microwave pulse (USMP) by Xing *et al.* successfully demonstrated a means of overcoming the conversion efficiency and resolution limited by the exciting pulse duration. The setup developed for experimental validation is schematically shown in Fig. 4(a). The core of the system is a USMP generator that is based on high-voltage-pulsed power technology. The USMP generator makes use of a charging circuit and Tesla

Table 1. Comparison of microwave sources by research groups.

Research group	Pulsed width	Main frequency	Producing mode	Peak power
R. A. Kruger	0.5 μs /1.0 ms ^{1–3}	434 MHz	HD	25 KW
V. Ntziachristos	10 ns ³⁴	~100 MHz	HD	~70 MW
	10/15/45 ns ³⁶	~400 MHz	HD	2.5/3.3/5.8 MW
D. Xing	10 ns ^{88–91}	434 MHz	HD	4–40 MW
	0.5 μs ^{79–82}	1.2 GHz	MM	300 KW
	0.3–1.1 μs ^{83–87}	6 GHz	MM	350 KW
L. V. Wang	0.6–2.2 μs ¹⁴	9.4 GHz	MM	10 KW
	0.5 μs ^{5,7,8}	3 GHz	MM	2 KW
	0.5 μs ⁶	9 GHz	MM	25/10 KW
H. Jiang	0.75 μs ^{39–41}	3 GHz	MM	70–100 KW
H. Xin	400 ns ⁴⁴	2.7–3.1 GHz	MC	5.2 KW
	0.5–1.5 μs ^{45–49}	3/1.4/2.5 GHz	MM	4/10/-KW
A. Arbabian	NA	2/2.1 GHz ^{50,61}	MC	120 W
S. K. Patch	700 ns ^{31,32}	108 MHz	MM	20 KW
Y. Zheng	NA	2.7/2 9/3.1 GHz ^{52–54}	MC	NA

Note: HD: High-voltage discharged mode; MM: Magnetron modulation mode; MC: Modulated continuous mode; NA: Not available.

Fig. 2. (a) A typical microwave generator working in the waveguide mode.⁹² (b) Several pulse widths of this microwave generator.⁸⁶

Fig. 3. Schematic diagram of high-voltage discharged mode.

transformer to produce high-voltage pulses and then discharge to the antenna for radiating microwave. The USMP generator has features of short pulse width, high peak power, simple structure and easy miniaturization. The pulse duration is

measured to be 10 ns with 800 ps rise time, and the peak output power is tunable from 2 to 40 MW, which satisfies the need of TAI.

To investigate the spatial resolution of the USMP-induced TAI, the cross-sections of two

Fig. 4. (a) Schematic diagram of the experimental setup of USMP TAI. (b) The comparison of microwave pulses of the USMP and the microsecond-level TAI. (c) The microwave pulse of the USMP.⁸⁸

Fig. 4. (Continued)

copper wires were imaged, as shown in Fig. 5. In the reconstruction image in Fig. 5(a), the two wires can be clearly detected. Figure 5(b) plots the normalized intensity profile of the cross-section image, and

the spatial resolution is determined to be $105\ \mu\text{m}$ according to the Rayleigh criterion. More importantly, the TA conversion efficiency was compared with short- and long-pulse (microsecond-level)

Fig. 5. Spatial resolution of the USMP TAI system. (a) TAI of two copper wires (0.13 mm in diameter). The inset figure at the bottom-right corner shows the diameter of the wire. (b) Line profile of the reconstructed image shown in (a) with $y = 1.84\ \text{mm}$. (c) The simulated peak-to-peak amplitude of TAI signal of the USMP and the microsecond-level TAI. (d) The experimental results of TAI signal amplitude.⁸⁸

excitation. Figure 5(c) gives the simulation results of the amplitude of the TAI signal, which was excited by various microwave pulse durations with identical pulse energy. Also, the experimental results in Fig. 5(d) show the amplitude of the TAI signal obtained by using a 10 ns microwave, which is 86-fold greater than that generated by a microwave pulse with duration of 450 ns. Therefore, it is expected that the enhancement of TA excitation efficiency should be about two orders of magnitudes.

However, some disadvantages still exist. One is that the frequency of the microwave limits to low frequency (100–900 MHz) and wide bandwidth (~ 100 MHz). At low frequency, microwave absorption coefficient is relatively small, which may be beneficial to imaging depth, but result in low SNR. Meanwhile, due to poor directivity of antenna, the energy scatters in whole space, resulting in a lower SNR, limiting its biomedical application.

3.3. Modulated continuous mode

The above two TAI modes adopt pulsed microwave, and the main frequency and pulse width are difficult

to change. Using modulated continuous microwave can solve the problem. As is shown in Fig. 6, the microwave generator is usually composed of signal source, modulator, power amplifier (PA) and antenna. The signal source can generate a low-power signal with variable frequency, and then the signal is modulated to pulsed mode by modulator. After PA, the high-power pulse transforms into unbounded medium propagation by antenna. Because the frequency is changed by signal source and modulator, the main frequency and pulse width are continuously adjustable. Therefore, this scheme has a wide range of applications, especially in material properties testing. However, the power of that microwave is restricted by PA, and its long-imaging time limits clinical application.

Zheng *et al.* used modulated continuous microwave to observe both TA and thermoacoustic resonance (TAR) effect, as shown in Fig. 7(a). A microwave generator (SMBV100A, Rohde & Schwarz) generates single and multi-pulse microwave source for the observation of TA and TAR effects, respectively. The microwave source is amplified up to 100 W with a microwave PA

Fig. 6. Schematic diagram of modulated continuous mode.

Fig. 7. (a) Schematic diagram of the experimental setup for the observation of both TA and TAR effect induced by single-pulse and multi-pulse microwave source. (b) The schematic of the experiment setup. The AWG sends out the stepped-frequency envelope signal that modulates the microwave carrier.⁵⁵

(ZHL-100 W-GAN+, Mini-Circuits). Considering the tissue absorption characteristics with respect to microwave frequencies, 440 MHz was chosen as the carrier frequency. The pulsed microwave signal is then fed into a custom-designed helical antenna operating at 440 MHz. By this system, they successfully generate detectable TAI signal.

Also, Arbabian used a stepped-frequency continuous-wave microwave for TAI, as shown in Fig. 7(b). The arbitrary wave generator (AWG) generates the low-frequency stepped-frequency envelope signal that then modulates a 2.1 GHz carrier. A gallium nitride PA is designed to bring the peak power up to 120 W (for pulse operation). Unlike other solutions that employ KW sources, the solution has a significantly lower peak power and is implemented completely with solid-state electronics. The pulse width is $1 \mu\text{s}$ and the microwave frequency is at 2.1 GHz. In the experiments, a planar immersion piezoelectric transducer with a central frequency of 0.5 MHz is used for detection. A low noise amplifier is employed for conditioning. After a low pass filter, the signal is averaged (typically 1000 times) and sampled by an oscilloscope. A linear stage setup is used to perform a B-Scan to construct the image of the sample.

4. Data Acquisition System

The equipment of ultrasound imaging technology, such as ultrasonic transducer, can be used in the detection of TAI signals, so that TAI has the similar resolution of ultrasound imaging. At present, the

ultrasonic transducers are made of piezoelectric crystal materials, polyvinylidene fluoride (PVDF) film and piezoelectric ceramic. Piezoelectric crystal materials can provide simple, accurate and high sensitivity of the signal detection. PVDF thin film material has the advantages of low acoustic impedance and frequency bandwidth, but its sensitivity is relatively low. Piezoelectric ceramic material has high sensitivity, but the bandwidth is narrow. The common acoustic detectors are divided into single-element and multi-element detectors, according to the number of elements. In accordance with the different detectors, the detection system is generally divided into single-element and multi-element detection systems, as shown in Table 2.

4.1. Scanning mode

The circular-scanning (CS) and linear-scanning (LS) modes are usually used in traditional TAI systems, as shown in Fig. 8. The CS mode was employed in most of the research on TAI techniques, which requires a single-element transducer rotating around the sample in circular fashion to acquire complete information of the target. A typical work flow of the single-element TAI system is shown in Fig. 9. The tissue absorbs microwave energy and then generates TA signals, and the single-element detector is placed at a distance from the center of the sample to receive the TAI signal. The received signal is amplified to the oscilloscope (which can also be replaced by a data acquisition card). In the oscilloscope, the TA signal is averaged

Fig. 8. Schematic diagram of LS (a) and CS (b).

Table 2. Comparison of DAS by research groups.

Research group	Type	Shape	Array element	Center frequency	Resolution
R. A. Kruger	Multi-element ¹⁻³	NA	64	1.0 MHz	~1 mm
V. Ntziachristos	Single-element ³⁴	NA	1	7.5 MHz	170 μm
	Single-element ³⁶	NA	1	3.5/15 MHz	45 μm
D. Xing	Single-element ⁸²⁻⁸⁵	NA	1	2.5/3.5 MHz	0.5 mm
		Linear ⁹⁷	128	2.5 MHz	2.2 mm
	Multi-element	Full ring ⁹¹	384	10 MHz	NA
		Flexible ⁹⁴	64	7.5 MHz	NA
L. V. Wang	Single-element ^{4,7,8}	NA	1	1/3.5 MHz	1.5 mm
	Multi-element ^{5,6}	Linear	30	2.25 MHz	1.9–2.5 mm
H. Jiang	Single-element ³⁹⁻⁴¹	NA	1	2.25 MHz	500 μm
H. Xin	Single-element ⁴⁴	NA	1	1 MHz	500 μm
	Multi-element ⁴⁵⁻⁴⁹	Linear	128	2.25 MHz	mm order
A. Arbabian	Single-element ^{50,51}	NA	NA	0.5 MHz	NA
S. K Patch	Single-element ³¹	NA	1	2.25 MHz	NA
	Multi-element ^{32,33}	Linear	96	1–4 MHz	250 μm
Y. Zheng	Single-element ⁵²⁻⁶⁴	NA	1	NA	NA

Notes: NA: Note available; DAS: Data acquisition system.

Fig. 9. Schematic diagram of single-element TAI system.⁸

multiple times, and then it is sent into the computer through the general interface bus. After the signal data transmission completes, the computer commands the stepper motor to drive the detector to rotate-scanning. When the detector rotates to the next position, the computer controls the acquisition system to scan signal for the next round until the 360° omnidirectional scanning is completed. In addition, a 2-D image can be reconstructed by using the filtered back-projection algorithm (or other algorithm), which can reflect microwave absorption

distribution of the biological tissue. Therefore, the single-element detector can acquire 2-D images by collecting signals using multi-position scanning (typically requiring dozens to hundreds of positions). Almost all groups above adopted single-element data acquisition system (DAS) at the beginning of the study, and details can be seen in Table 2.

Although this imaging mode can well reflect the electromagnetic wave absorption distribution of biological tissue, its experimental device and the source of the error are complex, and the data

acquisition time is long (typically 10–30 min), and the computational complexity of the reconstructed image takes a long time. Therefore, the development of multi-element detection systems is needed.

4.2. Multi-element data acquisition

As shown in Fig. 10(a), the process of multi-element data acquisition used by Xing *et al.* is depicted. In the imaging process, when the microwave pulses illuminate the object, the TA signals are detected by the 384 detectors simultaneously, and each group of 64 element signals passes through the switching system in sequence and is acquired by two 32-element data acquisition cards (NI5752, NI Inc., USA) at a sampling rate of 50 MHz. After switching six times, all 384 signals are acquired.

Finally, the 384-element signals are used to reconstruct a 2-D image via a filtered back-projection algorithm. Figures 10(b)–10(d) show images of the system. By the multi-element data acquisition system, the data acquisition time can reduce from dozens of minutes to 0.06 s, realizing real-time imaging. Figure 11 shows several typical multi-element detectors. Figure 11(a) is a commercial 128-element linear array transducer (SIUI Inc., L2L50A, Shantou, China). The central frequency of the transducer is 2.5 MHz with a bandwidth of 70%. Figure 11(b) shows that full ring transducer (10C384-1.62 8-R100 AHA001, Doppler Ltd, China) is made up of three parts, and every part contains 128 elements. The central frequency of the transducer is 10 MHz and the bandwidth ranges from 64.5% to 92.4%.

Fig. 10. (a) The process of signal acquisition of multi-element TAI system. (b) Photographs of the full ring transducer. (c) The 384-64 element switch system. (d) The 64-element data acquisition system.⁹¹

Fig. 11. Photographs of detectors. (a) Linear detector.⁹⁷ (b) Full ring detector.⁹¹ (c) Flexible detector.⁹⁴

However, the above ring detectors employ a fixed scanning radius to acquire information from different positions. For irregular samples, the scanning radius must be large enough to cover the longest axis of the imaging geometry. Thus, a large amount of coupling liquid required to fill in the gap between sample and detector hinders the practical application of TAI. Also, the linear detector has been widely used in medical ultrasonic imaging for body scanning, although it is not suitable for TAI of irregular objects due to image distortion and low utilization. Therefore, Xing *et al.* present a sample-clinging-scanning (SCS) model that combines the advantages of both ring and linear detector. To implement this model, a flexible multi-element transducer and adaptive back projection algorithm are presented. As shown in Fig. 11(c), most of the elements in the SCS model can be brought in close contact with sample due to the controllable shape of the transducer. Theoretically, it can adapt to samples with any shape without requiring coupling liquid, especially for breast cancer detection.

In summary, the single-element system is the most primitive equipment in TAI; it has the advantage of simple and cheap, but suffers from long-imaging time and instability, so it is suitable for material research. The multi-element system is specially designed for biomedical applications, mainly for breast detection. However, it applies only to single application. Among them, the linear detector, which has been widely used in medical ultrasonic imaging for body scanning, may become universal testing equipment after modification.

5. Biomedical Applications

Generally speaking, TAI can detect the microwave absorption anomaly at decimeter depth with sub-millimeter resolution, and the application of TAI needs to meet two basic conditions. One is the diseased tissue and the normal tissues have microwave absorption difference. For example, breast cancer tissue has absorption higher than normal breast tissue, and the ratio is estimated to be about 2:1–10:1. Another is the microwave energy can penetrate to the target, which can be guaranteed in humans, and the ultrasonic wave can spread out received by the detector. In particular, air and bone are against ultrasonic propagation, so anomaly within lung, human skull and bone is not appropriate to be detected.

5.1. Early breast cancer detection

The most important application of TAI is the early breast cancer detection. The main component of the breast is fat, which performs a low absorption in the microwave band. Therefore, in the early stages of the tumor, microwave absorption can increase significantly, due to ionic or water content growing up at the margin of the tumor. The change of absorption distribution of the tissues usually means physiological lesions, and the previous results suggest that the dielectric properties of malignant breast tissues are larger than the normal tissues. Therefore, the detection of breast cancer by TAI is feasible, and recently, more and more attention is paid to apply TAI to the detection of breast cancer, such as Kruger *et al.*, Wang *et al.* and Xing *et al.*

Among them, Xing *et al.* developed a system that provides a rapid and noninvasive approach for high-contrast breast cancer imaging. The results of TAI on the malignant breast tissue are shown in Fig. 12(a). In the results of TAI, a strong contrast was observed with satisfactory consistency in tumor position and shape, and the tumor-to-background contrast was estimated to be about 5.5:1, which demonstrates that TAI has a higher imaging contrast than X-ray mammography. Later, Xing *et al.* used the pathology model to view imaging capabilities. The pathology model is composed of excised human breast tumor and an adult ewe, as shown in Fig. 12(b). It can be seen that the difference between the normal tissue and the tumor in the TAI has a mean contrast of about 1:2.8. Further, Xing *et al.* performed a 3-D TAI experiment on breast model with a human tumor, and the results are shown in Fig. 12(c). Up to now, early breast cancer detection is the application that closest to the clinical, and the development direction is to clinical practice.

5.2. Prostate cancer

The prostate cancer has a higher microwave absorption than the surrounding normal tissues, so it is different on the TAI. Patch *et al.* presented a TAI system for performing *ex vivo* prostate cancer sample with very high frequency energy. The system utilized irradiation pulses of 700 ns duration exceeding 20 kW power and a dual-transducer system to achieve high contrast of prostate cancer. In the reconstruction images, the verumontanum can be

Fig. 12. The results of TAI of breast tumor. (a) Malignant breast tissue.⁸³ (b) The pathology model composed of excised human breast tumor and the excised breast of an adult ewe.⁹¹ (c) Three-dimensional TAI experiment on breast model.⁹³

clearly showed and conformed to histology slides in Fig. 13. Moreover, the values of TAI in select regions of the peripheral zone were associated with the disease status of the three histological specimens, and it can be seen that the strength of TA signal decreases with tumor involvement from specimens. Conclusion can be drawn that TAI can be a

potential detection method for prostate cancer in the future.

5.3. Foreign body detection

The normal biological tissue dominated by water has a high dielectric loss factor and absorbs microwave

Fig. 13. Mouse imaging experiments *ex vivo*. (a) Cross-sectional TAI acquired from an intact mouse inserted with a $230\ \mu\text{m}$ diameter copper wire. (b) The histological slice of the mouse made at approximately the same level. (c) The similar TAI acquired from the mouse without the presence of the copper wire.³⁵

Fig. 14. Comparison of foreign body detection between TAI and X-ray imaging. (a) X-ray imaging of the living mouse with the bamboo. (b) Photographs of the mouse and the bamboo strip. (c) Five TA images along the vertical direction of the bamboo strip.⁸⁴

strongly, whereas foreign bodies such as glass fiber, wood and bamboo with much less water content have much lower dielectric loss factors than the surrounding normal tissue. Therefore, TAI based on the differences of their electromagnetic properties has the potential to differentiate and image foreign body lesions. A TAI system by Xing *et al.* was developed to detect foreign targets in small animals. Figure 14(a) shows the X-ray imaging of the living mouse with the bamboo hidden in the body as marked in the dashed circle, and Fig. 14(b) shows the photographs of the inserted traumatic area of the mouse and the bamboo strip, respectively. Figure 14(c) shows the five TAI results along the vertical direction of the bamboo strip. Comparison with X-ray imaging and ultrasound imaging, the TAI can fill the blind zone for the X-ray band and may become an ideal and low cost modality for

radiolucent foreign body detection and imaging in humans and animals. In addition, Arbabian and co-workers⁹⁹ have achieved remote interrogation of highly dispersive media for deeply embedded abnormalities in 2015. This performance could enable a myriad of applications including medical imaging applications, security screening and nondestructive testing.

5.4. Small animal imaging

Besides cancer imaging and foreign body detection, TAI is also suitable for small animal imaging. Figure 15 shows the TAI results of a mouse reported by Ntziachristos *et al.* Figure 15(a) depicts reconstructed images from the mouse of the inserted line, and Fig. 15(b) depicts a photograph of the mouse slice obtained from the mice after frozen and cry

Fig. 15. (a–c) Three pathological sections of verumontanum. (d–f) The corresponding TAI reconstructions.³¹

Fig. 16. (a) The TAI result of index finger image of female volunteer. (b) The TAI result by setting the threshold of (a). (c) The photograph of the index finger. (d) The recovered relative microwave absorption in red line. (e) The MRI results of the whole hand. (e) and (f) The TAI results corresponding to (e).⁴³

sliced. It can be seen from the results that the copper wire on the left side of the spinal cord can be clearly detected. For comparison, Fig. 15(c) is obtained from the same mouse but without copper wire; therefore, it further shows anatomical details such as the spinal cord, the skin, the hindlimb and the urinary bladder, correlating well with the histological slice, such as in Fig. 15(b).

5.5. Joint imaging and brain imaging

As microwave absorption differences between the joint cavity and bone, joint imaging of TAI is possible. Jiang *et al.* proposed a TAI method to explore joint imaging, as shown in Fig. 16. The experimental results showed that the authors successfully detected joints and bones with TAI by demonstrating feasibility with three volunteers. Besides, Jiang *et al.* took an initial step to study the

feasibility of TAI toward imaging human brain. Several phantom experiments were conducted to demonstrate that it is indeed possible to use TAI for noninvasive imaging of hemorrhagic stroke with high contrast. These two applications represent the latest research in TAI. Of particular note is that the brain imaging in humans is difficult due to the thick skull, and the problem has to be solved in the future.

6. Prospects and Conclusions

In summary, under the efforts of the researchers, the excitation source, data acquisition system and applications of TAI have been intensively studied. It is believed that TAI has great potential applications in both biomedical research and clinical study. The most possible research directions in the future are the following:

1. *Expand the application range*: TAI has made great progress in imaging of breast cancer. Besides, TAI on prostate cancer has been reported. It indicates that TAI has certain advantages in cancer detection because the cancer tissue has relatively strong microwave absorption; therefore, expanding the application range to other cancers is rationale. Meanwhile, the examples of joint and brain imaging indicate that TAI has the potential to imaging damage of soft tissue, promising to be a new noninvasive functional imaging method.
2. *Develop multi-modal imaging*: TAI can map microwave absorption coefficient of biological tissues. However, the single-parameter imaging lacks other information and cannot meet the needs of getting more information of the body, which limits its availability. Therefore, in order to make doctors understand the internal structure of the body more accurately and comprehensively, combining TAI with other imaging modes (priority PAI and ultrasonic imaging) to recover the shortcomings of existing technology has practical significance and is urgently needed.

Acknowledgments

We thank Yi Luo, Baohua Qin, Yaping Li, Xin Wang, Yan Zhao and Wanling Luo for collecting references. This research is supported by the National Natural Science Foundation of China (61627827, 61331001, 81630046, 91539127), the Science and Technology Planning Project of Guangdong Province, China (2015B020233016, 2014B020215003, 2014A020215031, 2017A020215135), the Distinguished Young Teacher Project in Higher Education of Guangdong, China (YQ2015049), the Science and Technology Youth Talent for Special Project of Guangdong, China (2015TQ01X882), Young Teachers Scientific Research Cultivating Fund of South China Normal University (16KJ05) and China Postdoctoral Science Foundation (2017M610533).

References

1. R. A. Kruger, D. R. Reinecke, G. A. Kruger, "Thermoacoustic computed tomography — Technical considerations," *Med. Phys.* **26**, 1832–1837 (1999).
2. R. A. Kruger, K. D. Miller, H. E. Reynolds, W. L. Kiser, D. R. Reinecke, G. A. Kruger, "Breast cancer *in vivo*: Contrast enhancement with thermoacoustic CT at 434 MHz — Feasibility study," *Radiology* **216**, 279–283 (2000).
3. R. A. Kruger, W. L. Kiser, D. R. Reinecke, G. A. Kruger, K. D. Miller, "Thermoacoustic molecular imaging of small animals," *Mol. Imag.* **2**, 113–123 (2003).
4. L. V. Wang, X. Zhao, H. Sun, G. Ku, "Microwave-induced acoustic imaging of biological tissues," *Rev. Sci. Instrum.* **70**, 3744–3748 (1999).
5. G. Ku, L. V. Wang, "Scanning thermoacoustic tomography in biological tissue," *Med. Phys.* **27**, 1195–1202 (2000).
6. D. Feng, Y. Xu, G. Ku, L. V. Wang, "Microwave-induced thermoacoustic tomography: Reconstruction by synthetic aperture," *Med. Phys.* **28**, 2427–2431 (2001).
7. G. Ku, L. V. Wang, "Scanning microwave-induced thermoacoustic tomography: Signal, resolution, and contrast," *Med. Phys.* **28**, 4–10 (2001).
8. M. Xu, G. Ku, L. V. Wang, "Microwave-induced thermoacoustic tomography using multi-sector scanning," *Med. Phys.* **28**, 1958–1963 (2001).
9. Y. Xu, L. V. Wang, "Signal processing in scanning thermoacoustic tomography in biological tissues," *Med. Phys.* **28**, 1519–1524 (2001).
10. M. Xu, L. V. Wang, "Pulsed-microwave-induced thermoacoustic tomography: Filtered backprojection in a circular measurement configuration," *Med. Phys.* **29**, 1661–1669 (2002).
11. M. Xu, L. V. Wang, "Time-domain reconstruction for thermoacoustic tomography in a spherical geometry," *IEEE Trans. Med. Imag.* **21**, 814–822 (2002).
12. Y. Xu, D. Feng, L. V. Wang, "Exact frequency-domain reconstruction for thermoacoustic tomography. I. Planar geometry," *IEEE Trans. Med. Imag.* **21**, 823–828 (2002).
13. Y. Xu, M. Xu, L. V. Wang, "Exact frequency-domain reconstruction for thermoacoustic tomography. II. Cylindrical geometry," *IEEE Trans. Med. Imag.* **21**, 829–833 (2002).
14. M. Xu, Y. Xu, L. V. Wang, "Time-domain reconstruction algorithms and numerical simulations for thermoacoustic tomography in various geometries," *IEEE Trans. Biomed. Eng.* **50**, 1086–1099 (2003).
15. Y. Xu, L. V. Wang, "Effects of acoustic heterogeneity in breast thermoacoustic tomography," *IEEE Trans. Ultrason. Ferroelectr. Freq. Control* **50**, 1134–1146 (2003).
16. Y. Xu, L. V. Wang, G. Ambartsoumian, P. Kuchment, "Reconstructions in limited-view thermoacoustic tomography," *Med. Phys.* **31**, 724–733 (2004).

17. G. Ku, B. D. Fornage, X. Jin, M. Xu, K. K. Hunt, L. V. Wang, "Thermoacoustic and photoacoustic tomography of thick biological tissues toward breast imaging," *Technol. Cancer Res. Treat.* **4**, 559–565 (2005).
18. X. Jin, Y. Xu, L. V. Wang, Y. R. Fang, C. I. Zanelli, S. M. Howard, "Imaging of high-intensity focused ultrasound-induced lesions in soft biological tissue using thermoacoustic tomography," *Med. Phys.* **32**, 5–11 (2005).
19. X. Jin, L. V. Wang, "Thermoacoustic tomography with correction for acoustic speed variations," *Phys. Med. Biol.* **51**, 6437–6488 (2006).
20. Y. Xu, L. V. Wang, "Rhesus monkey brain imaging through intact skull with thermoacoustic tomography," *IEEE Trans. Ultrason. Ferroelectr. Freq. Control* **53**, 542–548 (2006).
21. X. Jin, C. Li, L. V. Wang, "Effects of acoustic heterogeneities on transcranial brain imaging with microwave-induced thermoacoustic tomography," *Med. Phys.* **35**, 3205–3214 (2008).
22. C. Li, M. Pramanik, G. Ku, L. V. Wang, "Image distortion in thermoacoustic tomography caused by microwave diffraction," *Phys. Rev. E* **77**, 031923 (2008).
23. M. Pramanik, G. Ku, C. Li, L. V. Wang, "Design and evaluation of a novel breast cancer detection system combining both thermoacoustic (TA) and photoacoustic (PA) tomography," *Med. Phys.* **35**, 2218–2223 (2008).
24. Y. Xie, B. Guo, J. Li, G. Ku, L. V. Wang, "Adaptive and robust methods of reconstruction (ARMOR) for thermoacoustic tomography," *IEEE Trans. Biomed. Eng.* **55**, 2741–2752 (2008).
25. M. Pramanik, G. Ku, L. V. Wang, "Tangential resolution improvement in thermoacoustic and photoacoustic tomography using a negative acoustic lens," *J. Biomed. Opt.* **14**, 024–028 (2009).
26. Z. Liu, L. Liu, Y. Xu, L. V. Wang, "Transcranial thermoacoustic tomography: A comparison of two imaging algorithms," *IEEE Trans. Med. Imag.* **32**, 289–294 (2013).
27. Y. He, C. Liu, L. Lin, L. V. Wang, "Comparative effects of linearly and circularly polarized illumination on microwave induced thermoacoustic tomography," *IEEE Antenn. Wirel. Propag. Lett.* (2016). doi: 10.1109/LAWP.2017.2652853.
28. S. K. Patch, "Thermoacoustic tomography — Consistency conditions and the partial scan problem," *Phys. Med. Biol.* **49**, 2305–2315 (2004).
29. A. T. Eckhart, R. T. Balmer, W. A. See, S. K. Patch, "Ex vivo thermoacoustic imaging over large fields of view with 108 MHz irradiation," *IEEE Trans. Biomed. Eng.* **58**, 2238–2246 (2011).
30. D. Fallon, L. Yan, G. W. Hanson, S. K. Patch, "RF testbed for thermoacoustic tomography," *Rev. Sci. Instrum.* **80**, 064301 (2009).
31. S. K. Patch, D. Hull, M. Thomas, S. K. Griep, K. Jacobssohn, W. A. See, "Thermoacoustic contrast of prostate cancer due to heating by very high frequency irradiation," *Phys. Med. Biol.* **60**, 689–708 (2015).
32. S. K. Patch, D. Hull, W. A. See, G. W. Hanson, "Toward quantitative whole organ thermoacoustics with a clinical array plus one very low-frequency channel applied to prostate cancer imaging," *IEEE Trans. Ultrason. Ferroelectr. Freq. Control* **63**, 245–255 (2016).
33. S. K. Patch, M. K. Covo, A. Jackson, "Thermoacoustic range verification using a clinical ultrasound array provides perfectly co-registered overlay of the Bragg peak onto an ultrasound image," *Phys. Med. Biol.* **61**, 5621 (2016).
34. D. Razansky, S. Kellnberger, V. Ntziachristos, "Near-field radiofrequency thermoacoustic tomography with impulse excitation," *Med. Phys.* **37**, 4602–4607 (2010).
35. S. Kellnberger, A. Hajiaboli, D. Razansky, V. Ntziachristos, "Near-field thermoacoustic tomography of small animals," *Phys. Med. Biol.* **56**, 3433–3444 (2011).
36. M. Omar, S. Kellnberger, G. Sergiadis, D. Razansky, V. Ntziachristos, "Near-field thermoacoustic imaging with transmission line pulsers," *Med. Phys.* **39**, 4460–4466 (2012).
37. A. Hajiaboli, S. Kellnberger, V. Ntziachristos, D. Razansky, "Design and time-domain analysis of a high-voltage impulsed test-bed for near-field thermoacoustic tomography," *Prog. Electromagn. Res.* **139**, 105–119 (2013).
38. L. Yao, G. Guo, H. Jiang, "Quantitative microwave-induced thermoacoustic tomography," *Med. Phys.* **37**, 3752–3759 (2010).
39. L. Huang, L. Yao, L. Liu, R. Jian, H. Jiang, "Quantitative thermoacoustic tomography: Recovery of conductivity maps of heterogeneous media," *Appl. Phys. Lett.* **101**, 244106 (2012).
40. L. Huang, J. Rong, L. Yao, W. Qi, D. Wu, J. Xu, H. Jiang, "Quantitative thermoacoustic tomography for ex vivo imaging conductivity of breast tissue," *Chin. Phys. Lett.* **30**, 124301 (2013).
41. Z. Chi, Y. Zhao, L. Huang, Z. Zheng, H. Jiang, "Thermoacoustic imaging of rabbit knee joints," *Med. Phys.* **43**, 6226–6233 (2016).
42. L. Huang, W. Cai, Y. Zhao, "In vivo tumor detection with combined MR-photoacoustic-thermoacoustic imaging," *J. Innov. Opt. Health Sci.* **9**, 1650015 (2016).

43. L. Huang, W. Qi, Y. Zhao, Z. Chi, N. Zhang, J. Rong, D. Lai, "Thermoacoustic imaging of human finger joints and bones," *X Acoust. Imaging Sens.* **1**, 28–31 (2015).
44. D. R. Bauer, X. Wang, J. Vollin, H. Xin, R. S. Witte, "Spectroscopic thermoacoustic imaging of water and fat composition," *Appl. Phys. Lett.* **101**, 033705 (2012).
45. X. Wang, D. R. Bauer, J. L. Vollin, D. G. Manzi, R. S. Witte, H. Xin, "Impact of microwave pulses on thermoacoustic imaging applications," *IEEE Antenn. Wirel. Propag. Lett.* **11**, 1634–1637 (2012).
46. X. Wang, D. R. Bauer, R. Witte, H. Xin, "Microwave-induced thermoacoustic imaging model for potential breast cancer detection," *IEEE Trans. Biomed. Eng.* **59**, 2782–2791 (2012).
47. T. Qin, X. Wang, Y. Qin, P. Ingram, G. Wan, R. S. Witte, H. Xin, "Experimental validation of a numerical model for thermoacoustic imaging applications," *IEEE Antenn. Wirel. Propag. Lett.* **14**, 1235–1238 (2015).
48. T. Qin, X. Wang, Y. Qin, G. Wan, R. S. Witte, H. Xin, "Quality improvement of thermoacoustic imaging based on compressive sensing," *IEEE Antenn. Wirel. Propag. Lett.* **14**, 1200–1203 (2015).
49. X. Wang, T. Qin, R. S. Witte, H. Xin, "Computational feasibility study of contrast-enhanced thermoacoustic imaging for breast cancer detection using realistic numerical breast phantoms," *IEEE Trans. Microw. Theory Tech.* **63**, 1489–1501 (2015).
50. H. Nan, A. Arbabian, "Stepped-frequency continuous-wave microwave-induced thermoacoustic imaging," *Appl. Phys. Lett.* **104**, 224104 (2014).
51. M. Aliroteh, H. Nan, A. Arbabian, "Microwave-induced thermoacoustic tomography for subcutaneous vascular imaging," *IEEE Ultrason. Symp. (IUS)*, pp. 1948–5727 (2016). DOI: 10.1109/ULTSYM.2016.7728643.
52. X. Feng, F. Gao, Y. Zheng, "Magnetically mediated thermoacoustic imaging toward deeper penetration," *Appl. Phys. Lett.* **103**, 083704 (2013).
53. X. Feng, F. Gao, Y. Zheng, "Modulatable magnetically mediated thermoacoustic imaging with magnetic nanoparticles," *Appl. Phys. Lett.* **106**, 153702 (2015).
54. G. Fei, X. Feng, Y. Zheng, "Advanced photoacoustic and thermoacoustic sensing and imaging beyond pulsed absorption contrast," *J. Opt.* **18**, 074006 (2016).
55. F. Gao, Y. Zheng, X. Feng, C. Ohl, "Thermoacoustic resonance effect and circuit modelling of biological tissue," *Appl. Phys. Lett.* **102**, 063702 (2013).
56. G. Zangerl, O. Scherzer, M. Haltmeier, "Circular integrating detectors in photo and thermoacoustic tomography," *Inverse Probl. Sci. Eng.* **17**, 133–142 (2009).
57. M. Agranovsky, P. Kuchment, "Uniqueness of reconstruction and an inversion procedure for thermoacoustic and photoacoustic tomography with variable sound speed," *Inverse Probl.* **23**, 2089–2102 (2007).
58. M. Haltmeier, O. Scherzer, P. Burgholzer, R. Nuster, G. Paltauf, "Thermoacoustic tomography and the circular radon transform: Exact inversion formula," *Math. Mod. Meth. Appl. Sci.* **17**, 635–655 (2007).
59. K. H. Lim, J. H. Lee, Q. H. Liu, "Thermoacoustic tomography forward modeling with the spectral element method," *Med. Phys.* **35**, 4–12 (2008).
60. A. Mashal, J. H. Booske, S. C. Hagness, "Toward contrast-enhanced microwave-induced thermoacoustic imaging of breast cancer: An experimental study of the effects of microbubbles on simple thermoacoustic targets," *Phys. Med. Biol.* **54**, 641–650 (2009).
61. D. Passeri, U. Sassi, A. Bettucci, E. Tamburri, F. Toschi, S. Orlanducci, M. L. Terranova, M. Rossi, "Thermoacoustic emission from carbon nanotubes imaged by atomic force microscopy," *Adv. Funct. Mater.* **22**, 2956–2963 (2012).
62. N. A. Rejesh, H. Pullagurla, M. Pramanik, M. Pramanik, "Deconvolution-based deblurring of reconstructed images in photoacoustic/thermoacoustic tomography," *J. Opt. Soc. Am. A* **30**, 1994–2001 (2013).
63. M. Pramanik, "Improving tangential resolution with a modified delay-and-sum reconstruction algorithm in photoacoustic and thermoacoustic tomography," *J. Opt. Soc. Am. A* **31**, 621–627 (2014).
64. O. Ogunlade, P. Beard, "Exogenous contrast agents for thermoacoustic imaging: An investigation into the underlying sources of contrast," *Med. Phys.* **42**, 170–181 (2015).
65. C. Zhang, Y. Wang, "A reconstruction algorithm for thermoacoustic tomography with compensation for acoustic speed heterogeneity," *Phys. Med. Biol.* **53**, 4971–4982 (2008).
66. X. Zhu, Z. Zhao, J. Wang, G. Chen, Q. Liu, "Active adjoint modeling method in microwave induced thermoacoustic tomography for breast tumor," *IEEE Trans. Biomed. Eng.* **61**, 1957–1966 (2014).
67. H. Peng, J. Hu, C. Hu, "Fe₃O₄@ MnO₂@ PAA nanoparticles for magnetically targeted microwave-thermal therapy guided by thermoacoustic imaging," *J. Biomater. Tissue Eng.* **6**, 12–19 (2016).
68. A. Dewantari, S. Y. Jeon, S. Kim, "Analysis of microwave-induced thermoacoustic signal generation using computer simulation," *J. Electromagn. Eng. Sci.* **16**, 1–6 (2016).

69. E. Hernández-Rosales, E. Cedeño, J. Hernandez-Wong, "Thermoacoustic and thermorefectance imaging of biased integrated circuits: Voltage and temperature maps," *Appl. Phys. Lett.* **109**, 041902 (2016).
70. S. Liu, Z. Zhao, X. Zhu, Z.-L. Wang, J. Song, B. Wang, Y.-B. Gong, Z.-P. Nie, Q. H. Liu, "Analysis of short pulse impacting on microwave induced thermo-acoustic tomography," *Prog. Electromagn. Res. C* **61**, 37–46 (2016).
71. B. M. Mihiretie, D. Cederkrantz, M. Sundin, "Thermal depth profiling of materials for defect detection using hot disk technique," *AIP Adv.* **6**, 085217 (2016).
72. Y. He, Y. Shen, C. Liu, "Suppressing excitation effects in microwave induced thermoacoustic tomography by multi-view Hilbert transformation," *Appl. Phys. Lett.* **110**, 053701 (2017).
73. S. Islam, S. Mahmud, M. Biglarbegian, "Effect of magnetic field on the Nusselt number of a multi-plate thermoacoustic system," *Int. J. Therm. Sci.* **108**, 145–158 (2016).
74. Y. Wu, Z. Tang, Y. Chi, "A simultaneous multi-probe detection label-free optical-resolution photoacoustic microscopy technique based on microcavity transducer," *J. Innov. Opt. Health Sci.* **6**, 1350027 (2013).
75. Y. Zhao, S. Yang, "Photoacoustic viscoelasticity imaging of biological tissues with intensity-modulated continuous-wave laser," *J. Innov. Opt. Health Sci.* **6**, 1350033 (2013).
76. J. Zhong, S. Yang, "Contrast-enhanced photoacoustic imaging using indocyanine green-containing nanoparticles," *J. Innov. Opt. Health Sci.* **7**, 1350029 (2014).
77. B. Yan, H. Qin, "Indocyanine green loaded graphene oxide for high-efficient photoacoustic tumor therapy," *J. Innov. Opt. Health Sci.* **9**, 1642001 (2016).
78. B. Yin, D. Xing, Y. Wang, Y. Zeng, Y. Tan, Q. Chen, "Fast photoacoustic imaging system based on 320-element linear transducer array," *Phys. Med. Biol.* **49**, 1339–1346 (2004).
79. D. Yang, D. Xing, H. Gu, Y. Tan, "Fast multi-element phase-controlled photoacoustic imaging based on limited-field filtered back projection algorithm," *Appl. Phys. Lett.* **87**, 194101 (2005).
80. L. Zeng, D. Xing, H. Gu, D. Yang, S. Yang, L. Xiang, "Fast microwave-induced thermoacoustic tomography based on multi-element phase-controlled focus technique," *Chin. Phys. Lett.* **23**, 1215–1218 (2006).
81. L. Zeng, D. Xing, H. Gu, D. Yang, S. Yang, L. Xiang, "High antinoise photoacoustic tomography based on a modified filtered backprojection algorithm with combination wavelet," *Med. Phys.* **34**, 556–563 (2007).
82. L. Nie, D. Xing, D. Yang, L. Zeng, Q. Zhou, "Detection of foreign body using fast thermoacoustic tomography with a multielement linear transducer array," *Appl. Phys. Lett.* **90**, 174109 (2007).
83. L. Nie, D. Xing, Q. Zhou, D. Yang, H. Guo, "Microwave-induced thermoacoustic scanning CT for high-contrast and noninvasive breast cancer imaging," *Med. Phys.* **35**, 4026–4032 (2008).
84. L. Nie, D. Xing, S. Yang, "In vivo detection and imaging of low-density foreign body with microwave-induced thermoacoustic tomography," *Med. Phys.* **36**, 3429–3437 (2009).
85. C. Lou, D. Xing, "Temperature monitoring utilizing thermoacoustic signals during pulsed microwave thermotherapy: A feasibility study," *Int. J. Hypertherm.* **26**, 338–346 (2010).
86. C. Lou, L. Nie, D. Xu, "Effect of excitation pulse width on thermoacoustic signal characteristics and the corresponding algorithm for optimization of imaging resolution," *J. Appl. Phys.* **110**, 083101 (2011).
87. C. Cao, L. Nie, C. Lou, D. Xing, "The feasibility of using microwave-induced thermoacoustic tomography for detection and evaluation of renal calculi," *Phys. Med. Biol.* **55**, 5203–5212 (2010).
88. C. Lou, S. Yang, Z. Ji, Q. Chen, D. Xing, "Ultrashort microwave-induced thermoacoustic imaging: A breakthrough in excitation efficiency and spatial resolution," *Phys. Rev. Lett.* **109**, 218101 (2012).
89. Y. Fu, Z. Ji, W. Ding, F. Ye, C. Lou, "Thermoacoustic imaging over large field of view for three-dimensional breast tumor localization: A phantom study," *Med. Phys.* **41**, 110701 (2014).
90. W. Ding, C. Lou, J. Qiu, Z. Zhao, Q. Zhou, M. Liang, S. Yang, Z. Ji, D. Xing, "Targeted Fe-filled carbon nanotube as a multifunctional contrast agent for thermoacoustic and magnetic resonance imaging of tumor in living mice," *Nanomed. Nanotechnol. Biol. Med.* **12**, 235–244 (2016).
91. F. Ye, Z. Ji, W. Ding, C. Lou, S. Yang, D. Xing, "Ultrashort microwave-pumped real-time thermoacoustic breast tumor imaging system," *IEEE Trans. Med. Imag.* **35**, 839–844 (2016).
92. W. Ding, Z. Ji, F. Ye, C. Lou, D. Xing, "Near-field microwave distribution measurement with a point detector base on thermoacoustic effect," *IEEE Trans. Microw. Theory Tech.* **63**, 3272–3276 (2015).
93. Z. Ji, C. Lou, S. Yang, D. Xing, "Three-dimensional thermoacoustic imaging for early breast cancer detection," *Med. Phys.* **39**, 6738–6744 (2012).

94. Z. Ji, W. Ding, F. Ye, C. Lou, D. Xing, "Shape-adapting thermoacoustic imaging system based on flexible multi-element transducer," *Appl. Phys. Lett.* **107**, 094104 (2015).
95. Z. Ji, C. Lou, Y. Shi, W. Ding, S. Yang, D. Xing, "A microwave detection way by electromagnetic and elastic resonance: Breaking the bottleneck of spatial resolution in microwave imaging," *Appl. Phys. Lett.* **107**, 164103 (2015).
96. Z. Ji, W. Ding, S. Yang, Q. Chen, D. Xing, "Remote measurement of microwave distribution based on optical detection," *Appl. Phys. Lett.* **108**, 014104 (2016).
97. Z. Ji, W. Ding, F. Hao, C. Lou, "Handheld thermoacoustic scanning system based on a linear-array transducer," *Ultrason. Imag.* **38**, 276–284 (2016)
98. Z. Ji, Y. Fu, S. Yang, "Microwave-induced thermoacoustic imaging for early breast cancer detection," *J. Innov. Opt. Health Sci.* **06**, 1350001 (2013).
99. H. Nan, K. C. Boyle, N. Apte, M. S. Aliroth, A. Bhuyan, A. Nikoozadeh, B. T. Khuri-Yakub, A. Arbabian, "Non-contact thermoacoustic detection of embedded targets using airborne-capacitive micromachined ultrasonic transducers," *Appl. Phys. Lett.* **106**, 084101 (2015).